


# THERMOPAR

## MG0 INSULATED



1	2	3	5	4	5	7	8	9
<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>

MIT

CALIBRATION	
J, JJ, K, KK, T, TT, E, EE	

1

WIRE LENGTH	
EX: 36"	

6

DIAMETER	
.063	1/16"
.125	1/8"
.188	3/16"
.250	1/4"

2

WIRE TYPE	
F	FIBERGLASS
S	FIBERGLASS W/ SSOB
A	FIBERGLASS W/ ARMOR
T	TEFLON
P	PVC

7

SHEATH MATERIAL	
S	316 SS
I	INCONEL 600

3

TERMINATION	
A	SPLIT LEADS
B	SPLIT LEADS W/ SPADES
C	STANDARD MALE CONN.
D	MINI MALE CONN.
E	STANDARD FEMALE CONN.
F	MINI FEMALE CONN.

8

SHEATH LENGTH	
EX: 6 1/2"	

4

JUNCTION	
G	GROUNDED
U	UNGROUNDED
E	EXPOSED

5

TRANSITION	
S	SPRING
0	NO SRPING

9